



Tri-Temp Bipolar Temperature Controller

Model # HCT-30

Ver. 1.0
3.2017



Information in this document is subject to change without notice.

No part of this document may be reproduced or transmitted without the express written permission of ALA Scientific Instruments Inc.

While every precaution has been taken in the preparation of this document, the publisher and the author assume no responsibility for errors or omissions, or for damages resulting from the use of information contained in this document. In no event shall the publisher and the author be liable for any loss of profit or any other commercial damage caused or alleged to have been caused directly or indirectly by this document.

© 2017 ALA Scientific Instruments, Inc. All rights reserved.

Printed: March 2017

ALA Scientific Instruments, Inc.
60 Marine Street
Farmingdale, NY 11735

Phone: +1 631.393.6401

Fax: +1 631.393.6407

Support@alascience.com

sales@alascience.com

www.alascience.com

Table of Contents:	Page #
SAFETY	0
1.0 INTRODUCTION	1
2.0 CONTROLLER	2
2.1 Front Panel	2
2.1.1 Power Switch (A)	2
2.1.2 Sensor B input (B)	2
2.1.3 Sensor A input and power out to channel A controlled device (C)	2
2.1.4 External Sensor Input (D)	2
2.1.5 Command in (E)	2
2.1.6 Output B (F)	2
2.1.7 Sensor A (G)	2
2.1.8 Display (H)	2
2.2 Rear Panel	3
2.2.1 Power IN	3
2.2.2 Fuse	3
3.0 PACKING LIST	4
4.0 SETUP	4
4.1 AC Power Connection	4
4.2 Assemble TRI-TEMP	4
5.0 MODES OF OPERATION	5
5.1 Temperature Mode	5
5.1.0 Heat Mode	6
5.1.1 Cool Mode	6
5.1.2 Heat/Cool Mode	6
5.2 Speed Mode	6
5.3 Voltage Mode	7
5.4 Control Feedback Mode	7
5.4.1 CTRL A	7
5.4.2 CTRL B (ACM Mode)	7
5.4.3 EXTERNAL Sensor Mode	8
5.5 Set Temperature Mode	8
5.5.1 Manual Set	8
5.5.2 External Set	9
5.5.3 Save Data	9
6.0 OUTPUT TEMPERATURE MONITOR	9
7.0 ERROR MESSAGES (ALARMS)	10
7.1 Out of Range	10
7.2 ACM Temperature Difference	10
8.0 POWER SUPPLY CHANNEL 3	11
9.0 HEAT/COOL	11
9.1 TRI-TEMP and Resistive Heating Devices	11
9.1.1 In-Line Heaters - HPC	11
9.1.2 Heated Stages - HCS	12
9.2 Peltier Devices	12
10.0 USING THE HEAT/COOL FEATURE	13

11.0 SPECIFICATIONS.....	14
12.0 CABLE PIN OUTS.....	14
13.0 WARRANTY.....	15

Safety

Do not use the device in a manner which is inconsistent with its intended function.

Do not expose the device to extreme conditions of humidity, hot or cold.

Do not use in a wet area or allow the power supply to become wet or submerged.

Keep ventilation holes on the rear of the unit clear from blockage at all times.

Do not use if wires are damaged or conductors are exposed.

Monitor feedback sensors to be sure unit stays in parameters and heated devices do not overheat.

This device has no clinical application and may not be used with human subjects at all.

1.0 Introduction

The Tri-Temp temperature controller was specifically designed for the demanding needs of the laboratory. Specifically built for use in electrophysiology it has features that make it a must for all types of labs where precision control of heated and/or cooled small apparatus is necessary. The TRI-TEMP is a 2.5 channel controller in that it can control two heating/cooling circuits and has a 2-12 V power output to power a third item without temperature sensor feedback. The system is designed to use 2252 ohm thermistors. In each control channel one thermistor is used to monitor the heating or cooling block (the component that does the heating or cooling itself), and the other sensor can be placed in the specimen. For example: The A sensor is built into the ALA HCPC (Heated/cooled perfusion cube) and the B sensor is placed in the recording bath. The control point for the temperature can be either sensor. A unique feature of the Tri-Temp is that it can sense if the B sensor has been removed from the bath chamber, or if the chamber has run dry. If that happens, the control point will automatically switch to the A sensor located within the heating/cooling block when a +/- 7 °C temperature difference is detected, and the block will not over heat or freeze as would happen if proper closed loop feedback is lost. Thus the specimen as well as the equipment is protected from damage.

The Tri-Temp is a heating or heating/cooling temperature controller. This means that it can operate a resistive device for heating, or a thermoelectric which can heat or cool. The power output is the same for whatever device you operate, but the polarity of the output voltage will switch when changing from heating to cooling. In a typical thermoelectric device, when power is applied, one side gets hot while the other side gets cold. If you want to cool an object, the side that gets cold is connected to the object that you want to cool. When the power is run for that configuration, the object will be cooled. The hot side of the thermoelectric must be connected to a heat sink to remove the excess heat. If you decide to heat the object, the voltage will be reversed and the thermoelectric will operate in reverse and the side exposed to the object will heat up and the other side will get cold.

The Tri-Temp features a membrane display and 5 modes of operation which are simply selected with the up and down arrows on the front panel. Power is supplied by a 15V universal power supply. Components that are to be plugged in must have a 9 pin DIN connector. The B thermistor needs a 2mm standard mono jack. Accurate control is achieved via PID control. The TRI-TEMP allows the user to select three speeds of operation, the slowest will result in the least overshoot of set point, the fastest the most overshoot. In the ACM (Automatic Control Mode), an alarm will sound, and the control point will be switched to the A sensor automatically if the difference in temperature between the set-point and the B sensor is more than 7°C. Please note that this feature is default OFF, and must be initiated each time the unit is turned on. The alarm will stop when the sensor is restored to the bath and the temperature stabilizes to within 7°C, or after 5 seconds, but the set point will be defaulted to A so the user must re-initialize the set point to B.

2.0 Controller

2.1 Front Panel



2.1.1 Power Switch (A)

Rocker switch turns unit on and off.

2.1.2 Sensor B input (B)

Plug in for B sensor, usually used to monitor a cell bath temperature.

2.1.3 Sensor A input and power out to channel A controlled device (C)

Din 8 connector with screw on shield that connects the A sensor, usually located within the heating or cooling device, to the Tri-Temp. The two bottom pins of the DIN output voltage to the heating or cooling device. Input calibrated for 2225 ohm thermistors.

2.1.4 External Sensor Input (D)

BNC connector I/O. Connect external sensor feedback signal (1mv/°C) from a different device as the control point for the TRI-TEMP.

2.1.5 Command in (E)

Input BNC used to adjust the set temperature of the TRI-TEMP. This function is only enabled when the EXT mode is selected. Input is 10mv/°C for entire temperature range of 0°C to 65°C.

2.1.6 Output B (F)

BNC output of sensor B thermistor reading. Output is calibrated to 10mv/°C across the temperature range of the TRI-TEMP. This signal can be sent to a data acquisition system or chart recorder.

2.1.7 Sensor A (G)

BNC output of Sensor A thermistor reading. Output is calibrated to 10mv/°C across the temperature range of the TRI-TEMP. This signal can be sent to a data acquisition system or chart recorder.

2.1.8 Display (H)

Displays all modes of the Tri-Temp, and all parameters including temperature readings of all inputs except Remote Temperature input.

2.1.9 Mode Keys (I)

Use up and down arrow keys to scroll through mode settings and to set temperature, power levels, speed of response, parameters, and memory. Press mode button to select mode.

2.1.10 USB connection (J)

Implementation pending

2.1.11 Overload 1 (K)

If channel 1 draws more than 2 amps, an alarm will sound a steady tone and this LED will light red. Press the button to re-set the channel. If the overload situation has not been resolved, the alarm will sound again.

2.1.12 Power on/off for channel 3 (L)

Press to turn on or off channel three power output.

2.1.13 Power Output Channel control knob (M)

Adjust this knob to set the power output of the banana plug connection. Power is 2-12 V up to 2A DC.

2.1.14 Reset channel 3 (N)

If an overload occurs a steady alarm tone will sound. Press to re-set. If condition is not cleared, alarm will sound again.

2.1.15 Channel 3 power output (O)

Red (+) and Black (-) banana plugs for power output. 2-12V DC 2A.

2.1.16 Reset for Channel 3 (P)

If an overload occurs a steady alarm tone will sound. Press to re-set. If condition is not cleared, alarm will sound again.

2.1.17 Grounding lug for chassis ground (Q)

Use standard banana plug to connect to chassis ground. This can be helpful to reduce electrical noise in e-phys recordings.

2.2 Rear Panel**2.2.1 Power IN**

Remote power supply input. (15 VDC @ 6A) 1mm center pin (+)

2.2.2 Fuse

TRI-TEMP, fuse housing. Fuse Specification: (6 Amp 5x20mm Slow Blow)

2.2.3 Ventilation

The rear panel has a fan that operates whenever the unit is powered up. Do not block or impinge the fan or the air holes. This will cause the TRI-TEMP to overheat and fail.

3.0 Packing List

The TRI-TEMP system comes with the following:

- TRI-TEMP Controller
- Universal Power Supply – 15V 6A DC Output
- Power Cord
- Instruction Manual

Carefully unpack the system from the shipping box and account for all parts listed above. Inspect all components for any signs of damage that might have occurred during shipping.

If there is damage to any component or a component is missing, immediately contact ALA Scientific Instruments.

If there is physical damage to any component of the system, do not dispose of any shipping materials (cardboard box, padding material). This will allow for a claim to be placed with the shipping company (UPS, FedEx).

4.0 Setup

The TRI-TEMP is designed to control ALA Scientific's Heat and/or Cool devices such as the HCPC, HPC, HCMIS and others. It is important to connect the device to the TRI-TEMP before turning TRI-TEMP ON.

4.1 AC Power Connection

The TRI-TEMP is designed to use a universal DC output power supply. Input AC voltage can vary from 100 VAC to 240 VAC 50/60 HZ.

- Use the supplied AC three prong power cord and connect it to a wall outlet near where the TRI-TEMP will be located.
- Plug power cable into universal power supply,
- Ensure that all cables are placed using GLP.
- If any cable is damaged (cut or damaged insulation) do not use. Replace with new cable.
- Connect universal power supply output connector into the power connector on the rear panel of the TRI-TEMP.

4.2 Assemble TRI-TEMP

The TRI-TEMP is a two channel bipolar temperature controller with second thermistor feedback ports, and DC power output channel. Using a second thermistor in the setup will allow for better monitoring of solution temperature in the bath. Adjustments to the TRI-TEMP can be more accurate with a second thermistor feedback.

The assembly process of the various temperature devices to the TRI-TEMP are as follows:

- Connect the device's 8-pin circular connector to the **Sensor A** connector on the TRI-TEMP front panel. Use TC4-CABLE (optional) for devices without a built-in cable.




- b. Connect thermistor TS-1/TS2 (optional) to **Sensor B** connector for bath temperature feedback.
- c. Place thermistor into cell chamber. The thermistor cable will need to be restrained in place to ensure bead remains completely submerged in the solution.
- d. Fill cell chamber to desired level making sure thermistor bead is completely submerged.
- e. Turn ON TRI-TEMP using rocker switch on the front panel.
- f. TRI-TEMP will power up.
- g. The TRI-TEMP displays will ask to:



Use the UP/DOWN buttons on the front panel to select YES or NO. On the first time powering up the TRI-TEMP there is no data stored other than the default values. The data stored will be the parameters from the last usage if you chose SAVE before power down.

Select either option. The default Screen will be displayed. The TRI-TEMP is ready to be programmed with your required parameters.



If previous parameters are retrieved they will appear on the screen. **Note that the power level always defaults to 2V no matter what it was set to before saving as a safety feature.**

5.0 Modes of Operation

The TRI-TEMP single channel bipolar temperature controller has many modes of operation to facilitate the control of the device connected. The *MODE* button is used to scroll through the various modes displayed. Once highlighted, the *UP/DOWN* buttons are used to scroll through each mode's options.



5.1 Temperature Mode

Heating and cooling can be set to heat, cool, or both. Press mode button until the heat is highlighted, as shown below. The default setting is heat when the TRI-TEMP is turned ON.

5.1.0 Heat Mode

The default Heat mode for the TRI-TEMP is *HEAT*. Upon powering up the TRI-TEMP for the first time, the *HEAT* display will be highlighted. The Heat mode is used with devices that are designed to be maintained above room temperature. Devices using resistive elements such as ALA Scientific’s In-line heater (HPC), objective heater (OBJHEATER), or stage heater (HCS) can be used in this mode.

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

5.1.1 Cool Mode

The *COOL* mode is the second option in the temperature mode display field. Use the UP button to scroll to the *COOL* mode from the *HEAT* mode. Only devices having thermo-electric based modules (Peltier) can be used in this mode. Such devices include ALA Scientific’s HCPC and HCMIS. Use the *COOL* mode for systems that need to be kept below room temperature.

A CTRL 22.0°C	COOL
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

5.1.2 Heat/Cool Mode

The H/C mode is used with thermo-electric based devices that will be actively heated or cooled. Although the UP/DOWN buttons can be used to change temperatures it is recommended that the external command feature be used. An input analog COMMAND signal can quickly change temperatures. (See more information on Heat/Cool below on page 18)

Note: *The Heat/Cool feature is only applicable to devices that can heat and cool. If the device being used only heats, the H/C feature should not be used. Serious damage to your device can occur. Furthermore, Peltier (thermoelectric) based heat/cool devices have voltage restrictions. For instance all ALA cooling devices must be limited to 7V or serious damage to the cooling/heating elements will occur.*

5.2 Speed Mode

The speed of response of the system can be adjusted with three levels; low, medium and high. Press the *MODE* button until the speed function with the three bars is highlighted. The default speed setting is high. Use the UP/DOWN buttons to change speed mode. The slow setting gives the most accuracy to limit overshoot, the high setting gives the fastest response, but the most overshoot.

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

Operational Note: Changing the speed setting modifies the PID algorithm that regulates the controller.

5.3 Voltage Mode

The voltage that the TRI-TEMP outputs to the connected device can be changed by pressing the *MODE* button until the *OUT* display is highlighted. Use the UP/DOWN buttons to change the voltage output. The voltage ranges from 2 to a maximum of 12 volts. Upon start up, default voltage is 2V to prevent damage to heating/cooling devices. For best results, heating and cooling, this setting should be set on the maximum voltage your device can handle.

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 12 V

Note: Be aware of the maximum voltage your device can tolerate. Read the specifications of the devices being used and do not exceed the maximum voltage permitted.

5.4 Control Feedback Mode

Feedback for temperature control can come from one of three sensor inputs. Thermistors connected to *SENSOR A*, *SENSOR B* or *EXT SENSOR BNC* port can be used as feedback.

Press the *MODE* button until both A and B are highlighted. Use the UP/DOWN buttons to change the input source.

A CTRL 22.0°C	HEAT
CH B X.X°C	SPEED ■ ■ ■
SET 20.0°C	OUT 2 V

Note: If Channel B is not being used (nothing is plugged in) then the B control will not work, and will appear as X.X°C.

5.4.1 CTRL A

The TRI-TEMP default temperature control feedback is *SENSOR A* (*A CTRL*). CH B can remain unused or a thermistor (TS-1M or TS-2M) can be connected and used to monitor the temperature in the cell bath or any other point of interest. Channel A will follow the set point requested.

A CTRL 37.0°C	HEAT
CH B 36.9°C	SPEED ■ ■ ■
SET 37.0°C	OUT 7 V

5.4.2 CTRL B (ACM Mode)

When *SENSOR B* is selected as the temperature control point, channel B will follow the value requested in the set point temperature.

Since *SENSOR B* is primarily used for measuring temperature in a cell bath, there is the possibility that it can fall out of the bath or that the bath can dry up, losing the feedback link with the TRI-TEMP. This may

B CTRL 37.0°C	HEAT
CHA ACM 37.1°C	SPEED ■ ■ ■
SET 37.0°C	OUT 7 V

cause your system to overheat as it tries to heat a dry bath while the sensor keeps reporting only room temperature. For this reason the TRI-TEMP is automatically set into ACM mode or Automatic Control Mode when *SENSOR B* is the control point. This mode is designed to protect your sample and equipment from overheating or freezing.

In ACM mode, if the temperature difference between A and B is more than 7°C the system will sound a pulsing audible alarm for about 5 seconds and automatically switch control from B, back to Sensor A (at the B set point level). Once the problem has been corrected, you can switch the control point back to B.

Always check the positioning of the thermistor and the solution level to assure proper feedback.

Note: Be aware that ACM is only available in HEAT, and COOL mode, it will not function in H/C Heat/Cool mode since driving that mode with an External input might force a more than 7° difference between A and B. Furthermore, since cooling below ambient is harder than heating above ambient, the ACM mode may not be usable in situations where you are trying to cool the bath near 0°C. A greater than 7° difference may occur and you will need to control from sensor A and just monitor B.


5.4.3 EXTERNAL Sensor Mode

The TRI-TEMP can operate with feedback from an external device as the control point. Devices that have a 1mv/°C output such as temperature meters can be used. This mode allows for the use of various thermocouples (PT-100). **The temperature from the external sensor will not be shown on the display of the TRI-TEMP.** Even though the display does not show the value, it will interpret the input and use it as the feedback point. Use the external device to read the temperature.




To activate this mode, use the UP/DOWN buttons to select *EX*. Connect your external monitor to the front panel BNC labeled *EXT SENSOR*.

Note: ACM mode is not enabled in this mode. Take care to monitor sensor placement and solution levels.




5.5 Set Temperature Mode

The TRI-TEMP has 2 ways to change the temperature set point; manual and external modes. To enter the set point temperature mode, press the *MODE* button until *SET* is highlighted. The TRI-TEMP default set point is 20°C @ 2V.




5.5.1 Manual Set

The TRI-TEMP temperature set point can be adjusted manually from the controller front panel. With the *SET* point mode highlighted, press the *MODE* button once to manually adjust the temperature set point. Use the UP/DOWN buttons to move to the desired temperature set point. A maximum temperature of 64.9°C can be set in manual mode.




5.5.2 External Set

The TRI-TEMP temperature set point can be controlled via an external analog voltage. To change to **EX** control mode with the **SET** mode highlighted, press the UP arrow button until EX is displayed.



Connect a BNC cable from an analog source to the **COMMAND** input (CMD) on the TRI-TEMP. Command input maximum voltage is 5 volts. Each movement of 10mV will shift the control point 1°C to a maximum of 50°C.



5.5.3 Save Data

The TRI-TEMP will save the parameters set in the mode screens to enable faster set up later. As you scroll through the modes, the TRI-TEMP will ask whether to save the data or not.



Continue scrolling if to continue using the current parameters without saving them or press the DOWN button. Press the UP button to save the parameters for future use. Any changes made after this screen will have to be saved prior to turning the TRI-TEMP off.

Next time the TRI-TEMP is turned ON, the display will ask whether to retrieve the data saved. Press the UP button to load the saved parameters. Press the DOWN button to start from the default settings.




Note: For safety reasons the output voltage parameter is always defaulted to 2V.

6.0 Output temperature Monitor

The TRI-TEMP has analog voltage output ports for each sensor input. **OUTPUT A** and **OUTPUT B** BNC connectors can be used to monitor temperature via a data acquisition system.

Each 10mV output corresponds to a 1°C temperature reading. For example, a **SENSOR A** temperature reading of 45°C translates to a 0.45V (450mV) output.



7.0 Error Messages (Alarms)

The TRI-TEMP has audible alarms when certain parameters or feedback signals are outside of specifications. These alarms are meant to alert the user of improper connections and to protect the device from overheating. The overheat alarm and overload alarm is a continuous tone. It will sound when sensor A reads 65°C. *Please note that not all systems are equipped with over temperature alarms due to customer requirements.*

A CTRL 64.0°C	HEAT
CH B 64.0°C	SPEED ■ ■ ■
SET OVER	OUT 7V

The overload alarm can be silenced by pressing the reset button for the lighted channel. If the overload is cleared, the alarm will silence.



7.1 Out of Range

The range for the set point is from 0°C to 65°C. If the set point is set above or below this range, the input will not be accepted.

A CTRL 64.0°C	HEAT
CH B 64.0°C	SPEED ■ ■ ■
SET OVER	OUT 7V

7.2 ACM Temperature Difference

When *SENSOR B* is selected as the temperature control point, the TRI-TEMP will go into Automatic Control Mode (ACM mode).

An audible alarm will sound in ACM mode if the difference in temperature between *SENSOR A* and *SENSOR B* is greater than 7°C.

B CTRL 37.2°C	HEAT
CH A ACM 44.5°C	SPEED ■ ■ ■
SET 37.0°C	OUT 7V

The alarm will stop if the temperature difference falls below 7°C. If the temperature difference remains for more than 10 seconds, the TRI-TEMP will revert to *SENSOR A* as the control point and the alarm will silence. The alarm will pulse for about 5 seconds.

Check that *SENSOR B* is properly placed and solution levels are appropriate before resetting the TRI-TEMP. Use the MODE button to scroll back to the control feedback mode to reset the TRI-TEMP back to *SENSOR B* as the control point.

8.0 Power Supply Channel 3

Channel three of the TRI-TEMP is a power supply. The power goes from 0-12V and is set by the adjustment knob. Use the POWER Channel 3 button to toggle the power on and off. (LED will indicate that power is on) Power output is limited to 2A. If more current is drawn, the channel will give an audible alarm and turn off. You will need to press the reset button to continue to use the power output, but the short circuit or high power demand will need to be cleared or the channel will continue to alarm. The polarity of the output is shown by the colors of the banana plugs for the output. Red is positive (+) and black is negative (-). Polarity cannot be reversed internally.



9.0 Heat/Cool

The TRI-TEMP is a bipolar which means it can power heating or cooling a devices with thermoelectric modules. Thermoelectric modules can heat or cool depending on how the polarity of the DC power supplied to them is arranged. ALA Scientific has temperature devices that can be used to heat or cool various parameters within an experiment such as the cell solution, perfusate, stages, microscope objectives and much more.

9.1 TRI-TEMP and Resistive Heating Devices

The TRI-TEMP is designed to control various types of resistive element heaters. All of ALA Scientific's resistive element heating devices such as the HPC (Heated Perfusion Cube), HCS (Heated Chamber Stage), OBJHEAT (Objective lens Heater) and HEATING Pad (Heated silicone pad) are able to be controlled and monitored by the TRI-TEMP.

If a third party device is used with the TRI-TEMP it must follow certain criteria as to not damage the TRI-TEMP or not be properly controlled. The resistance of the device must not be below 5 ohms and the feedback thermistor must be a 2252 ohm @ 25°C type.

9.1.1 In-Line Heaters - HPC

The TRI-TEMP *HEAT* mode is designed to control and monitor in-line heaters such as ALA Scientific's Heated Perfusion Cube (HPC).

The HPC is a device for heating up a perfusion flow to a cell bath. It is typically used for a flow rate of up to 5ml/min.

Suppose you want it to heat a 2ml/min flow into a 36°C 1.5ml volume cell bath.

Follow the setup instructions (section 4.0) to assemble the TRI-TEMP system.




Connect the 8-pin cable from the HPC to the *SENSOR A* circular connector on the TRI-TEMP.

When the unit is powered ON, **HEAT** and **high SPEED** will be the default settings displayed. Do not change these settings. Proceed to change the *OUT* voltage to 10 or higher. The HPC works best at 12V.

A CTRL	39.1°C	HEAT
CH B	37.1°C	SPEED ■ ■ ■
SET	39.0°C	OUT 12 V

If a TS-1M or TS-2M (optional) is available, connect it to *SENSOR B* and enable the display of CH B.

Once all the desired functions are set, go to the *SET* mode and either chose the manual or EXTERNAL mode to adjust the temperature set point. In manual mode, use the UP/DOWN buttons to change the **set point (SET)** to the desired temperature.



Depending on the flow rate, once the internal cube temperature, as read on *Sensor A* is stable, it may be possible to switch control to the *SENSOR B* which can be placed in the cell bath for more accurate temperature control. Once the *SENSOR B* is positioned, control can be switched to the *SENSOR B*. The ACM mode will prevent more than a 7° difference between the two sensors. If it is necessary to heat the HPC much hotter than the desired temperature in the bath (high flow rate), switch off the ACM mode and adjust the set point for A to a point where the B sensor reads the desired temperature. Often a flowing bath perfusion will need to be heated more than 10° hotter than the desired bath temperature because there is so much heat loss. Just note that in these circumstances the ACM mode and thus *SENSOR B* as the control point cannot be used.

9.1.2 Heated Stages – HCS

It is often necessary to heat the stage housing a cell chamber in order to maintain a unified temperature control of the experiment. The HCS connects to the HCT controller with the TC4-CABLE (optional).

The HCS has a steel outer ring allowing for magnetic tools to be used. The inner ring contains the heater and feedback sensor.

The TRI-TEMP parameters are set to Heat only and it is recommended to use the maximum voltage (12V) in the *OUT* setting.



9.2 Peltier Devices

The TRI-TEMP is a bipolar temperature controller. It is well suited to control ALA Scientific's Heated Cooled Perfusion Cube (HCPC) and the HCMIS micro incubator stage.



Setup for a Peltier device:


- Connect the Peltier device to the TRI-TEMP.
- Turn ON TRI-TEMP.
- Change the temperature mode to H/C mode.
- Select the SPEED mode. Fast is recommended.

- Change the *OUT* voltage to 7V or less. **The maximum voltage permitted before damage may occur to the thermoelectric module is 7V.**
- The temperature feedback is already defaulted to *SENSOR A*. At this point if *SENSOR B* is also used change the temperature feedback mode to also read *CH B*. place the *SENSOR B* thermistor into the cell bath and secure in place.
- Once all the desired functions are set, go to the *SET* mode and either chose the manual or *EXT*ernal mode to adjust the temperature set point. In manual mode, use the *UP/DOWN* buttons to change the **set point (SET)** to the desired temperature.
- Both the *HCPC* and the *HCMIS* peltier devices have water cooled heat sinks to help remove heat generated during cooling. This helps in reaching lower temperatures. When cooling make sure water or coolant is flowing through the heat sink.

Note: Be sure to check the maximum voltage whenever using thermoelectric devices as these are usually lower than heat-only devices with resistive heaters.

10.0 Using the Heat/Cool Feature

Certain devices can both heat and cool, like ALA's Heating/Cooling Micro-Incubation System (*HCMIS*) or Heated/Cooled Perfusion Cube (*HCPC*).



Heat/Cool mode is useful for moving the set point temperature above and below room temperature, or doing temperature controlled experiments where the set point must move up and down the temperature scale rapidly: i.e faster than normal decay to ambient temperature. This mode is particularly useful with an external command driving the temperature change in an experiment where changing temperature is the variable.

To set up the heat cool feature:

- Turn the controller on.
- Press the mode button until the heat section is highlighted (heat is default).
- Press the up/down buttons to change it to H/C (heat/cool) mode.
- Set the *SPEED* of the system. If the temperature oscillates too much, the speed may be too fast. If so, adjust the speed by pressing the *MODE* button until speed is highlighted and use the *UP/DOWN* buttons to adjust speed.
- Set the *OUT* voltage for the device. ALA Scientific's Peltier devices have a maximum voltage rating of 7V.
- Set the second thermistor (*CH B* if one is available) using the *MODE* button to highlight the control point section.
- The *SET* temperature can be manually adjusted or the *EX* mode can be used. The Heat/Cool mode may be desirable when commanding the HCT from an external source for creating increasing and decreasing temperature ramps.
- Check the performance to be sure that the system can follow the commands in real time. You may need to adjust ramp and soak times to accommodate the hardware's ability to react to the temperature controller.

Note: To simply lower the temperature below room temperature of about - 5°C or less, use the cool feature, not H/C mode.

11.0 Specifications

Maximum Output Voltage	12 VDC
Maximum Output Current	2.0 A
Minimum Load Resistance	2.0 Ω
Manual Voltage Range	2 to 12 VDC
Maximum Output Power	24 W @ 10 ohm load
Universal AC adapter	Input 100 – 240 VAC 50/60 Hz Output 15 VDC @ 6.0 A max output
Power Fuse (5 x 20 mm)	6 A Slow Blow 250V
Sensor Type	2252 ohm Thermistor
Front Panel Input/output	Chan 1&2 Sensor A – Thermistor (2252) DIN Chan 1&2 Sensor B (Input) – Thermistor (2252) 2mm mono jack Chan 1&2 Command Input – BNC 10mV/°C Chan 1&2 Output A- BNC 10mV/°C Chan 1&2 Output B- BNC 10mV/°C Chan 1&2 EXT SENSOR (input) – BNC 1mV/°C Chan 3 power output-4mm Banana, 0-12V DC
Temperature Range	0°C to 65°C
Enclosure (W x H x D)	48.5 x 9.0 x 28 cm (19 x 3.5 x 11 in)
Enclosure Weight	6 Lbs.


12.0 Cable Pin Outs

DIN Connector as seen from the front panel:

Pin 1 and 2 are for sensor.

Pins 1 & 3 are connected internally.

Pin 7 and 8 are power out



13.0 Warranty

ALA Scientific Instruments, Inc. agrees to warranty this product against defects in material and workmanship for one year from date of shipment. Remedy shall be limited to replacement or repair of the item(s) at ALA's discretion. The usage of this product by the user will indicate the user's understanding of the use of this product as set forth in this manual. Neither ALA Scientific Instruments, Inc., nor any of its affiliates will be held responsible for damage to laboratory equipment, including microscopes, resulting from the use or misuse of this product, including damage resulting from inputs exceeding specified limits that result in malfunction to or from this device. The user asserts that he/she is aware of the electrical output and that he/she will insure that it does not exceed manufacturers' recommendation for heat/cool devices used in conjunction with the HCT.

In the event that warranty repairs are necessary, shipping charges to the factory are the customer's responsibility. Return charges will be paid by ALA Scientific Instruments for warranty repairs only.

This instrument is not for clinical use. It is strictly for basic research in a laboratory setting. It has no clinical application whatsoever and cannot be used on human subjects.

Email: Support@alascience.com
Voice: 631 393-6401
Web: www.alascience.com

ALA Scientific Instruments, Inc.
60 Marine Street
Farmingdale NY 11735